

Test Certificate

with reference to OFI Test Report No. 2001972/12290

Client: Elbatrade HandelsgmbH
 z.H.: Hr. Richson
 Favoritner Gewerbering 15-17 / 1.OG
 1100 Wien

Order: Testing for disinfecting effect

Samples: UV- air disinfection device "LumeeLamp Sterilizer Dual NX"

Receipt of Samples: 2020-10-29

Test Method: The device was tested with an Escherichia phage MS2 (= bacteriophage). The bioaerosol was sprayed into the device using an aerosol generator and collected after discharge using an absolute filter. The replicable virus particles were evaluated microbiologically.
 Tests were carried out in the individual technical departments of OFI within the scope of competence of the authorised signatories according to OFI QM manual.

Test Results:

Tested sample mode	Disinfection effect against viruses Mean value and standard deviation
„turbo mode“	99,91 ± 0,02%
„silent mode“	99,88 ± 0,01%

Approved: G. Ettenber



Date: 12. 01. 2021

DI Gabriele Ettenberger
 Director in charge,
 Pharma, Medical Devices and Hygiene

This certificate comprises 1 page.

It may be made available to third parties, either free of charge or against payment, if the full wording and if OFI is expressly named as the author. All tests applied are subject to a quality assurance programme according to EN ISO/IEC 17025:2017.